



C.I. MAJESTIC UNIVERSAL S.A.S

**TECHNICAL SHEET MEDICAL COVERALLS
PROTECTION**

DATE

**SEPTEMBER 1
2020**

**COMMERCIAL NAME OF THE
PRODUCT**

MEDICAL COVERALLS PROTECTION



GENREAL DESCRIPTION AND MATERIALS

MATERIAL

FABRIC: 100% POLYESTER ANTI-FLUID.

FABRIC SPECIFICATIONS

TEXTILE ANTI FLUID IS A PROTECTIVE FINISH THAT REPELS LIQUIDS AND ACCIDENTAL SPLASHES. HAND WASHABLE (WATER AND SOAP OR SODIUM HYPOCHLORITE)

PACKAGING

INDIVIDUAL BAG

INDICATIONS

DURABILITY 3 MONTHS.
OFFERS FULL BODY PROTECTION, EASY TO PUT ON CLOTHING.
FRONT ZIPPERS, SLEEVES, AND BOOTS WITH ELASTIC CUFFS.
LIGHT, DURABLE AND STRONG.

USES

PROTECTION OF THE PHYSICIAN REPELLENT TO BODY FLUIDS AND OTHER THAN PRESENT DURING THE PROCEDURES.

DESIGN SPECIFICATIONS

**COLOR
SIZE**

FRONT ZIPPER, SLEEVES AND BOOTS WITH ELASTIC CUFFS.
LIGHT, DURABLE AND STRONG.
WITH LONG SLEEVES AND SPRING CUFFS.
FRONT ZIPPER, SLEEVES AND BOOTS WITH ELASTIC CUFFS.
ZIP POCKETS.
BLUE
S TO XL.

STANDARDIZATION

SURGICAL FIELDS, COATS AND SUITS

NTC 5623

SURGICAL FIELDS, COATS AND SUITS FOR CLEAN AIR FOR SURGICAL USE, AS MEDICAL DEVICES FOR PATIENTS, CLINICAL STAFF AND EQUIPMENT.

PART 1. GENERAL REQUIREMENTS FOR MANUFACTURERS PROCESSORS AND PRODUCTS.

NTC 5624 - SURGICAL FIELDS, TIES AND SUITS FOR CLEAN AIR FOR SURGICAL USE, AS MEDICAL DEVICES FOR PATIENTS, CLINICAL STAFF AND EQUIPMENT.

PART 2. TEST METHODS. NTC 5625 - SURGICAL FIELDS, COATS AND SUITS FOR CLEAN AIR FOR SURGICAL USE, AS MEDICAL DEVICES FOR PATIENTS, CLINICAL STAFF AND EQUIPMENT.

PART 3. REQUIREMENTS AND LEVELS OF OPERATION.
PRODUCT DESCRIPTION:

OF NON-WOVEN FABRIC, REPELLENT TO LIQUIDS, BUT PERMEABLE TO AIR AND OPTIMAL RESISTANCE TO BREAKAGE. REINFORCED IN SLEEVES AND FRONT.

THE SLEEVES ARE FINISHED WITH ELASTIC COTTON CUFFS, ADJUSTABLE TO THE WRIST AND ABOUT 8 TO 10 CM LENGTH. ROUND NECK, ANATOMICALLY ADJUSTABLE TO THE NECK.